

The American College of Radiology, with more than 30,000 members, is the principal organization of radiologists, radiation oncologists, and clinical medical physicists in the United States. The College is a nonprofit professional society whose primary purposes are to advance the science of radiology, improve radiologic services to the patient, study the socioeconomic aspects of the practice of radiology, and encourage continuing education for radiologists, radiation oncologists, medical physicists, and persons practicing in allied professional fields.

The American College of Radiology will periodically define new practice parameters and technical standards for radiologic practice to help advance the science of radiology and to improve the quality of service to patients throughout the United States. Existing practice parameters and technical standards will be reviewed for revision or renewal, as appropriate, on their fifth anniversary or sooner, if indicated.

Each practice parameter and technical standard, representing a policy statement by the College, has undergone a thorough consensus process in which it has been subjected to extensive review and approval. The practice parameters and technical standards recognize that the safe and effective use of diagnostic and therapeutic radiology requires specific training, skills, and techniques, as described in each document. Reproduction or modification of the published practice parameter and technical standard by those entities not providing these services is not authorized.

Revised 2021 (Resolution 37)\*

## **ACR–SPR PRACTICE PARAMETER FOR THE PERFORMANCE AND INTERPRETATION OF SKELETAL SURVEYS IN CHILDREN**

---

### **PREAMBLE**

This document is an educational tool designed to assist practitioners in providing appropriate radiologic care for patients. Practice Parameters and Technical Standards are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care<sup>1</sup>. For these reasons and those set forth below, the American College of Radiology and our collaborating medical specialty societies caution against the use of these documents in litigation in which the clinical decisions of a practitioner are called into question.

The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the physician or medical physicist in light of all the circumstances presented. Thus, an approach that differs from the practice parameters, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in the practice parameters when, in the reasonable judgment of the practitioner, such course of action is indicated by the condition of the patient, limitations of available resources, or advances in knowledge or technology subsequent to publication of the practice parameters. However, a practitioner who employs an approach substantially different from these practice parameters is advised to document in the patient record information sufficient to explain the approach taken.

The practice of medicine involves not only the science, but also the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment. Therefore, it should be recognized that adherence to these practice parameters will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The sole purpose of these practice parameters is to assist practitioners in achieving this objective.

---

<sup>1</sup> Iowa Medical Society and Iowa Society of Anesthesiologists v. Iowa Board of Nursing, 831 N.W.2d 826 (Iowa 2013) Iowa Supreme Court refuses to find that the *ACR Technical Standard for Management of the Use of Radiation in Fluoroscopic Procedures* (Revised 2008) sets a national standard for who may perform fluoroscopic procedures in light of the standard's stated purpose that ACR standards are educational tools and not intended to establish a legal standard of care. See also, Stanley v. McCarver, 63 P.3d 1076 (Ariz. App. 2003) where in a concurring opinion the Court stated that "published standards or guidelines of specialty medical organizations are useful in determining the duty owed or the standard of care applicable in a given situation" even though ACR standards themselves do not establish the standard of care.

## **I. INTRODUCTION**

This practice parameter was revised collaboratively by the American College of Radiology (ACR) and the Society for Pediatric Radiology (SPR).

A skeletal survey is a systematically performed series of radiographic images that encompasses the entire skeleton or those anatomic regions appropriate for the clinical indications. Radiographic skeletal surveys are used for a variety of clinical problems in children. Common clinical indications include suspected nonaccidental trauma/child abuse, skeletal dysplasia, metabolic disorder, or bony metastases [1-10]. The goal of the skeletal survey is to accurately identify focal and diffuse abnormalities of the skeleton, including acute or healing fractures, bone lesions, evidence of metabolic bone disease, or characteristics of skeletal dysplasia, and to differentiate them from developmental changes and other anatomic variants that may occur in infants and children.

## **II. INDICATIONS**

Indications for skeletal surveys include, but are not limited to:

1. Known or suspected physical abuse in infants and young children
2. Known or suspected skeletal dysplasias, syndromes, and metabolic disorders
3. Known or suspected neoplasia and related disorders

## **III. QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL**

See the [ACR–SPR Practice Parameter for General Radiography](#) [11].

### **A. Physician**

The radiologist should understand the utility of alternate imaging techniques such as ultrasonography, CT, nuclear medicine, and MRI in order to fulfill a consultative role and to interpret pediatric skeletal surveys in the context of other available imaging results.

### **B. Medical Physicist**

The medical physicist should have the training to aid in the development and routine review of imaging protocols in the various imaging modalities in pediatrics and provide consultative advice in the adoption of new technologies and routine testing of these systems. The physicist should also be able to give dose estimates and general advice on the use of such information to the radiologist when requested.

### **C. Technologist**

The technologist should have training and experience in performing radiographic examinations in infants and children. In particular, the technologist should be familiar with positioning and patient restraint, as well as customary measures to minimize radiation exposure. The technologist should be aware of the unique circumstances created when children with suspected abuse are brought to the radiology department by caretakers, guardians, and child protective service representatives.

## **IV. SPECIFICATIONS OF THE EXAMINATION**

The written or electronic request for radiographic skeletal surveys should provide sufficient information to demonstrate the medical necessity of the examination and allow for its proper performance and interpretation.

Documentation that satisfies medical necessity includes 1) signs and symptoms and/or 2) relevant history (including known diagnoses). Additional information regarding the specific reason for the examination or a provisional diagnosis would be helpful and may at times be needed to allow for the proper performance and interpretation of the examination.

The request for the examination must be originated by a physician or other appropriately licensed health care provider. The accompanying clinical information should be provided by a physician or other appropriately licensed

health care provider familiar with the patient's clinical problem or question and consistent with the state's scope of practice requirements. (ACR Resolution 35 adopted in 2006 – revised in 2016, Resolution 12-b)

The skeletal survey examination should be performed in accordance with traditional principles of high-quality diagnostic radiography. These include proper technique factors, positioning, collimation, image identification, and immobilization methods.

The imaging protocol for the skeletal survey will depend on the particular clinical indication. Additionally, the radiologist should consider modifying a complete protocol based on imaging that was already performed on the infant or child so as to minimize unnecessary radiation exposure.

#### A. Known or Suspected Nonaccidental Trauma/Child Abuse

Each anatomic region (see Table 1) should be imaged with a separate radiographic exposure to ensure uniform image density and maximize image sharpness. A single radiograph (babygram) of the entire infant should not be performed. Each extremity should be radiographed in at least the frontal projection. Radiographs of the axial skeleton should be obtained in two projections: anteroposterior (AP)/frontal and lateral. Right and left posterior oblique views of the entire rib cage aid analysis for rib fractures [12-17]. Additional views as needed should be obtained to fully document suspected abnormalities and may include lateral (orthogonal) views of the long bones [18], a Towne view of the skull, AP and lateral views of selected joints, or additional obliquities of the ribs or other areas of concern [19]. The skeletal survey should be reviewed by a qualified radiologist as defined in Section III. A second interpretation by a pediatric radiologist may add value [20].

A follow-up skeletal survey may be indicated in the setting of nonaccidental injury. Many times, a complete repeat examination is appropriate, though a limited follow-up examination could also be considered [21-26]. Postmortem skeletal surveys may also be helpful [27,28] in addition to postmortem histopathology.

Some clinical and radiological findings (such as burns, bruising, single unexplained fractures, and retinal and intracranial hemorrhages), particularly in children younger than 2 years old, suggest a need to consider child abuse and to perform a skeletal series [29-32].

In the skeletal series, evaluation for any traumatic injury should be performed. In particular, classic metaphyseal lesions, such as metaphyseal corner and buckle handle fractures that are considered highly specific for nonaccidental trauma/child abuse (as are posterior rib, scapular, spinous process and sternal fractures), are best seen by the skeletal series [10,33,34]. Knowledge of variants and simulators of traumatic findings are necessary [35,36].

Skeletal series findings may suggest complementary imaging for diagnosis, including CT/MR for brain analysis [37], nuclear scintigraphy [38,39], and extremity CT for a focal area of concern [40]. CT of the brain may use 3-D skull reconstructions at no additional radiation cost for complementary analysis of calvarium [41,42].

#### B. Skeletal Dysplasias, Syndromes, and Metabolic Disorders

##### 1. Skeletal dysplasias and syndromes

Imaging of skeletal dysplasias, including those in children with disproportionate stature and a wide variety of syndromes, should conform to the standard skeletal survey protocol (see the Complete Skeletal Survey Table below), with the following exceptions:

- a. Entire arms and legs can be exposed on a single film when the size of the child permits.
- b. In newborns and young infants, whole-body AP and lateral radiographs may be appropriate, but separate views of the skull (frontal and lateral), hands (posteroanterior (PA)), and feet (AP) are advisable. Lateral views of the feet and ankles may be useful in selected cases.
- c. As previously noted, review by a qualified physician is essential, with additional views obtained as required (eg, flexion and extension lateral views of the cervical spine for certain skeletal dysplasias).
- d. In some patients, selected images of specific regions or additional views may be appropriate, depending on the differential diagnoses being considered [43-46].

2. Metabolic disorders (rickets and rickets-like disorders)

In general, it is not necessary to survey the entire bony skeleton for metabolic disorders. A targeted examination focusing on the appropriate anatomic regions of interest to include PA views of the wrists and AP views of the knees is recommended. Occasionally, a complete skeletal survey may be warranted [47].

C. Neoplasia and Related Conditions

Langerhans cell histiocytosis can present with a solitary bone lesion or widely disseminated disease. A complete skeletal survey should be performed as part of the initial imaging evaluation. Additional orthogonal projections of areas suspected to be abnormal on clinical or other imaging grounds should be obtained. A complete skeletal survey may also be obtained as part of the evaluation for metastatic disease to the bone [48-50].

COMPLETE SKELETAL SURVEY TABLE

APPENDICULAR SKELETON
Right and left humerus (AP)
Right and left ulna & radius (AP)
Right and left hand (PA)
Right and left femur (AP)
Right and left tibia & fibula (AP)
Right and left foot (AP)

AXIAL SKELETON
Thorax (AP, lateral, right and left obliques), to include sternum, ribs [51,52], and thoracic and upper lumbar spine
Abdomen/pelvis, to include the thoracolumbar spine and sacrum (AP)
Lumbosacral spine (lateral)
Skull (frontal and lateral), to include cervical spine (if not completely visualized on lateral skull)*

**V. DOCUMENTATION**

Reporting should be in accordance with the [ACR Practice Parameter for Communication of Diagnostic Imaging Findings](#) [53].

An official interpretation (final report) of the examination should be included in the patient’s medical record. The report should provide a concise description of all sites of definite and suspected abnormality. A standardized summary with descriptive text may be helpful [54]. When a constellation of radiographic findings is sufficient to raise strong suspicion of abuse, this should be so stated in the radiology report and communicated to the referring physician, and this communication should be documented in the final report. A physician diagnosing suspected child abuse is often legally required to notify local child protection authorities. Thus, if the attending physician does not report the case, the radiologist may still be required to do so.

**VI. EQUIPMENT SPECIFICATIONS, RADIOGRAPHIC TECHNIQUE, AND RADIATION DOSE**

The quality of a skeletal system survey is a function of the resolution and dose efficiency of the imaging system. Attention to contrast and resolution should be addressed when selecting imaging equipment along with the protocols and processing algorithms used.

---

\*If a contemporaneous head CT was performed prior to the skeletal survey and its resolution sufficient to provide high quality multiplanar 2-D reconstructions and surface rendered 3-D skull reconstruction, the radiologist may exclude the AP and lateral views of the skull from the skeletal series.

Radiology departments should carefully select their digital radiographic systems with particular attention to high diagnostic efficiency. Appropriate technical factor and processing parameter optimization should be made for the demanding application of skeletal survey for suspected child abuse [3,9,55,56]. The lowest possible radiation dose consistent with acceptable diagnostic image quality should be used, particularly in pediatric examinations.

In infants, the entire examination should be performed with suitable high-detail imaging. A grid should not be used in an infant. In the toddler and older child, the increase in patient thickness compared with that of the infant may require greater radiation dosage for optimal penetration. Scatter due to increased thickness may reduce image contrast unless a moving grid is used to image larger body regions. Kilovoltage peak (kVp) should be set at a sufficiently low level to provide adequate subject contrast.

When selecting a digital radiographic system, it should have high spatial resolution and exhibit optimal dose efficiency characteristics. If the system has multiple-resolution mode capability, the high-resolution mode should be used. The higher-resolution mode may require an increase in milliamperere-seconds to maintain the signal-to-noise ratio and to optimize visualization of skeletal structures. Digital processing menus and image display parameters should be selected to enhance bone detail [9,55,57-64]. It should be noted that the use of such bone detail-processing algorithms will result in an increase in noise, which may require compensation by increasing milliamperere-seconds. Optimal use of high-resolution imaging systems will result in an increase in radiation dose compared with typical low-dose systems widely used for general pediatric imaging. When judiciously applied for appropriate indications, this increased dose is justified so as to obtain superior skeletal detail. When modern high-detail imaging systems are coupled with meticulous radiographic technique, the patient dose remains well within accepted levels, and the associated risks are extremely small. Appropriate collimation should be used to limit radiation exposure to the anatomic area of interest and reduce unnecessary scatter. Past versions of the skeletal series document have underlined the need to make sure shielding is not included within the imaging field so as to avoid the negative impacts of shielding on exposure controls and image quality as well as the obscuration of anatomy, compromising diagnosis. Current ACR, SPR, and AAPM documents have concluded that shielding should no longer be used [65].

The kVp range employed in skeletal survey imaging is 55 to 70, which is generally used for all images of the appendicular skeleton, skull, and spine of infants with the exception of images of the hands and feet, which may be performed at a lower kVp. In the toddler, the kVp is increased as necessary when imaging the skull and spine. The milliamperere-seconds is adjusted according to the kVp, image recording system, and x-ray equipment design (eg, filtration, generator, etc). Techniques should be governed more by patient size rather than absolute age. Filtration can be used to remove unwanted low energy x-rays that add to the overall examination dose but are unlikely to penetrate a larger body region. The source to image distance is 101.6 cm (40 in). Skeletal survey images in infants are usually performed on the tabletop. In toddlers and older children, dose considerations may require sacrificing some resolution and/or contrast. The use of the under-table cassette slot in conjunction with a moving antiscatter grid is likely to produce optimal results in larger patients. Meticulous positioning and collimation over each anatomic region are essential. Both joints are included in all long-bone images. Chest imaging uses bone detail technique.

Equipment performance monitoring should be in accordance with the [ACR-AAPM Technical Standard for Diagnostic Medical Physics Performance Monitoring of Radiographic Equipment](#) [66].

## VII. RADIATION SAFETY IN IMAGING

Radiologists, medical physicists, registered radiologist assistants, radiologic technologists, and all supervising physicians have a responsibility for safety in the workplace by keeping radiation exposure to staff, and to society as a whole, “as low as reasonably achievable” (ALARA) and to assure that radiation doses to individual patients are appropriate, taking into account the possible risk from radiation exposure and the diagnostic image quality necessary to achieve the clinical objective. All personnel that work with ionizing radiation must understand the key principles of occupational and public radiation protection (justification, optimization of protection and application of dose limits) and the principles of proper management of radiation dose to patients (justification, optimization and the use of dose reference levels)

[http://www-pub.iaea.org/MTCD/Publications/PDF/Pub1578\\_web-57265295.pdf](http://www-pub.iaea.org/MTCD/Publications/PDF/Pub1578_web-57265295.pdf)

Nationally developed guidelines, such as the ACR [Appropriateness Criteria®](#), should be used to help choose the most appropriate imaging procedures to prevent unwarranted radiation exposure.

Facilities should have and adhere to policies and procedures that require varying ionizing radiation examination protocols (plain radiography, fluoroscopy, interventional radiology, CT) to take into account patient body habitus (such as patient dimensions, weight, or body mass index) to optimize the relationship between minimal radiation dose and adequate image quality. Automated dose reduction technologies available on imaging equipment should be used whenever appropriate. If such technology is not available, appropriate manual techniques should be used.

Additional information regarding patient radiation safety in imaging is available at the Image Gently® for children ([www.imagegently.org](http://www.imagegently.org)) and Image Wisely® for adults ([www.imagewisely.org](http://www.imagewisely.org)) websites. These advocacy and awareness campaigns provide free educational materials for all stakeholders involved in imaging (patients, technologists, referring providers, medical physicists, and radiologists).

Radiation exposures or other dose indices should be measured and patient radiation dose estimated for representative examinations and types of patients by a Qualified Medical Physicist in accordance with the applicable ACR technical standards. Regular auditing of patient dose indices should be performed by comparing the facility's dose information with national benchmarks, such as the ACR Dose Index Registry, the NCRP Report No. 172, Reference Levels and Achievable Doses in Medical and Dental Imaging: Recommendations for the United States or the Conference of Radiation Control Program Director's National Evaluation of X-ray Trends. (ACR Resolution 17 adopted in 2006 – revised in 2009, 2013, Resolution 52).

## VIII. QUALITY CONTROL AND IMPROVEMENT, SAFETY, INFECTION CONTROL, AND PATIENT EDUCATION

Policies and procedures related to quality, patient education, infection control, and safety should be developed and implemented in accordance with the ACR Policy on Quality Control and Improvement, Safety, Infection Control, and Patient Education appearing under the heading *ACR Position Statement on Quality Control & Improvement, Safety, Infection Control, and Patient Education* on the ACR website (<https://www.acr.org/Advocacy-and-Economics/ACR-Position-Statements/Quality-Control-and-Improvement>).

## ACKNOWLEDGEMENTS

This practice parameter was revised according to the process described under the heading *The Process for Developing ACR Practice Parameters and Technical Standards* on the ACR website (<https://www.acr.org/Clinical-Resources/Practice-Parameters-and-Technical-Standards>) by the Committee on Practice Parameters – Pediatric Radiology of the ACR Commission Pediatric Radiology and the Committee on Practice Parameters – Body Imaging (Musculoskeletal) of the ACR Commission on Body Imaging, in collaboration with the SPR.

Writing Committee – members represent their societies in the initial and final revision of this practice parameter

### ACR

Harris L. Cohen, MD, FACR, Chair  
Stephen F. Miller, MD  
Terry Levin, MD  
Suzanne S. Long, MD  
Esben Vogelius, MD  
Thaddeus Wilson, PhD

### SPR

Siddharth Jadhav, MD  
Jessica Leschied, MD  
Andy Tsai, MD

### Committee on Practice Parameters – Pediatric Radiology

(ACR Committee responsible for sponsoring the draft through the process)

Terry L. Levin, MD, FACR, Chair

Jane Sun Kim, MD

### Committee on Practice Parameters – Pediatric Radiology

John B. Amodio, MD, FACR	Jennifer A Knight, MD
Jesse Berman, MD	Jessica Kurian, MD
Tara M. Catanzano, MB, BCh	Matthew P. Lungren, MD, MPH
Harris L. Cohen, MD, FACR	Helen R. Nadel, MD
Kassa Darge, MD, PhD	Erica Poletto, MD
Dorothy L. Gilbertson-Dahdal, MD	Richard B. Towbin, MD, FACR
Lauren P. Golding, MD	Andrew T. Trout, MD
Safwan S. Halabi, MD	Esben S. Vogelius, MD
Jason Higgins, DO	

### Committee on Body Imaging (Musculoskeletal)

(ACR Committee responsible for sponsoring the draft through the process)

Catherine C. Roberts, MD, Chair	Suzanne S. Long, MD
Jeffrey M. Brody, MD, FACR	Kambiz Motamedi, MD
Bethany U. Casagrande, DO	Carlos A. Rivera, BSc
Elaine S Gould, MD, FACR	Aleksandr Rozenberg, MD
Mary K. Jesse, MD	Naveen Subhas, MD
Kenneth S. Lee, MD	

Richard A. Barth, MD, FACR, Chair, Commission on Pediatric Radiology  
Andrew B. Rosenkrantz, MD, Chair, Commission on Body Imaging  
David B. Larson, MD, MBA, Chair, Commission on Quality and Safety  
Mary S. Newell, MD, FACR, Chair, Committee on Practice Parameters and Technical Standards

### Comment Reconciliation Committee

Amy Patel, MD, Chair	David B. Larson, MD, MBA
Adam Specht, MD, FACR, Co-Chair	Paul A. Larson, MD, FACR
Richard A. Barth, MD, FACR	Jessica Leschied, MD
Leslie Ann Bord, MD	Terry L. Levin, MD, FACR
Harris L. Cohen, MD, FACR	Suzanne S. Long, MD
Richard Duszak Jr., MD, FACR	Stephen F. Miller, MD
Dustin A. Gress, MS	Mary S. Newell, MD, FACR
Siddharth Jadhav, MD	Andy Tsai, MD
Jane Sun Kim, MD	Esben Vogelius, MD
Amy Kotsenas, MD, FACR	Thaddeus Wilson, PhD

## REFERENCES

1. Belfer RA, Klein BL, Orr L. Use of the skeletal survey in the evaluation of child maltreatment. *Am J Emerg Med* 2001;19:122-4.
2. Kellogg ND. Evaluation of suspected child physical abuse. *Pediatrics* 2007;119:1232-41.
3. Kleinman PL, Kleinman PK, Savageau JA. Suspected infant abuse: radiographic skeletal survey practices in pediatric health care facilities. *Radiology* 2004;233:477-85.
4. Merten DF, Radkowski MA, Leonidas JC. The abused child: a radiological reappraisal. *Radiology* 1983;146:377-81.
5. Offiah AC, Hall CM. Observational study of skeletal surveys in suspected non-accidental injury. *Clin Radiol* 2003;58:702-5.
6. Guillerman RP, Braverman RM, Parker BR. Imaging studies in the diagnosis and management of pediatric malignancies. In: Pizzo PA, Poplack DG, eds. *Principles and Practice of Pediatric Oncology*. 5th ed. Philadelphia, Pa: Lippincott Williams & Wilkins; 2006:236-89.



7. Lachman RS. *Taybi and Lachman's Radiology of Syndromes, Metabolic Disorders and Skeletal Dysplasias, 5th ed.* 5th ed. St. Louis: Elsevier-Mosby; 2007.
8. Lachman RS, Rappaport V. Fetal imaging in the skeletal dysplasias. *Clin Perinatol* 1990;17:703-22.
9. Barber I, Perez-Rossello JM, Wilson CR, Kleinman PK. The yield of high-detail radiographic skeletal surveys in suspected infant abuse. *Pediatr Radiol* 2015;45:69-80.
10. Kleinman P. *Diagnostic Imaging of Child Abuse.* 3 ed. Cambridge: Cambridge University Press; 2015.
11. American College of Radiology. ACR–SPR practice parameter for general radiography. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/RadGen.pdf>. Accessed January 14, 2020.
12. Hansen KK, Prince JS, Nixon GW. Oblique chest views as a routine part of skeletal surveys performed for possible physical abuse--is this practice worthwhile? *Child Abuse Negl* 2008;32:155-9.
13. Ingram J, Connell J, Hay T, Strain J, Mackenzie T. Oblique radiographs of the chest in nonaccidental trauma. *Emerg Radiol* 2000;7:42-46.
14. Marine MB, Corea D, Steenburg SD, et al. Is the new ACR-SPR practice guideline for addition of oblique views of the ribs to the skeletal survey for child abuse justified? *AJR Am J Roentgenol* 2014;202:868-71.
15. Lindberg DM, Harper NS, Laskey AL, Berger RP. Prevalence of abusive fractures of the hands, feet, spine, or pelvis on skeletal survey: perhaps "uncommon" is more common than suggested. *Pediatric emergency care* 2013;29:26-9.
16. Kleinman PK, Morris NB, Makris J, Moles RL, Kleinman PL. Yield of radiographic skeletal surveys for detection of hand, foot, and spine fractures in suspected child abuse. *AJR Am J Roentgenol* 2013;200:641-4.
17. Kemp AM, Butler A, Morris S, et al. Which radiological investigations should be performed to identify fractures in suspected child abuse? *Clin Radiol* 2006;61:723-36.
18. Karmazyn B, Duhn RD, Jennings SG, et al. Long bone fracture detection in suspected child abuse: contribution of lateral views. *Pediatr Radiol* 2011.
19. Phillips KL, Bastin ST, Davies-Payne D, et al. Radiographic skeletal survey for non-accidental injury: systematic review and development of a national New Zealand protocol. *Journal of medical imaging and radiation oncology* 2015;59:54-65.
20. Karmazyn B, Wanner MR, Marine MB, Tilmans L, Jennings SG, Hibbard RA. The added value of a second read by pediatric radiologists for outside skeletal surveys. *Pediatr Radiol* 2019;49:203-09.
21. Hansen KK, Keeshin BR, Flaherty E, et al. Sensitivity of the limited view follow-up skeletal survey. *Pediatrics* 2014;134:242-8.
22. Jha P, Stein-Wexler R, Coulter K, Seibert A, Li CS, Wootton-Gorges SL. Optimizing bone surveys performed for suspected non-accidental trauma with attention to maximizing diagnostic yield while minimizing radiation exposure: utility of pelvic and lateral radiographs. *Pediatr Radiol* 2013;43:668-72.
23. Harper NS, Eddleman S, Lindberg DM. The utility of follow-up skeletal surveys in child abuse. *Pediatrics* 2013;131:e672-8.
24. Singh R, Squires J, Fromkin JB, Berger RP. Assessing the use of follow-up skeletal surveys in children with suspected physical abuse. *The journal of trauma and acute care surgery* 2012;73:972-6.
25. Bennett BL, Chua MS, Care M, Kachelmeyer A, Mahabee-Gittens M. Retrospective review to determine the utility of follow-up skeletal surveys in child abuse evaluations when the initial skeletal survey is normal. *BMC research notes* 2011;4:354.
26. Zimmerman S, Makoroff K, Care M, Thomas A, Shapiro R. Utility of follow-up skeletal surveys in suspected child physical abuse evaluations. *Child Abuse Negl* 2005;29:1075-83.
27. Hughes-Roberts Y, Arthurs OJ, Moss H, Set PA. Post-mortem skeletal surveys in suspected non-accidental injury. *Clin Radiol* 2012;67:868-76.
28. McGraw EP, Pless JE, Pennington DJ, White SJ. Postmortem radiography after unexpected death in neonates, infants, and children: should imaging be routine? *AJR Am J Roentgenol* 2002;178:1517-21.
29. Fagen KE, Shalaby-Rana E, Jackson AM. Frequency of skeletal injuries in children with inflicted burns. *Pediatr Radiol* 2015;45:396-401.
30. Wood JN, Fakeye O, Mondestin V, Rubin DM, Localio R, Feudtner C. Development of hospital-based guidelines for skeletal survey in young children with bruises. *Pediatrics* 2015;135:e312-20.
31. Paine CW, Scribano PV, Localio R, Wood JN. Development of Guidelines for Skeletal Survey in Young Children With Intracranial Hemorrhage. *Pediatrics* 2016;137.
32. Mattheij M, Venstermans C, de Veuster I, et al. Retinal haemorrhages in a university hospital: not always abusive head injury. *Acta Neurol Belg* 2017;117:515-22.



33. Bittman M P-RJ, Tracy D, Shabani A, Lee E. Musculoskeletal traumatic disorders. *Pediatric Radiology: Practical Imaging Evaluation of Infants and Children* 2018;1189-240.
34. Raynor E, Konala P, Freemont A. The detection of significant fractures in suspected infant abuse. *J Forensic Leg Med* 2018;60:9-14.
35. Quigley AJ, Stafrace S. Skeletal survey normal variants, artefacts and commonly misinterpreted findings not to be confused with non-accidental injury. *Pediatr Radiol* 2014;44:82-93; quiz 79-81.
36. Tsai A, Johnston PR, Perez-Rossello JM, Breen MA, Kleinman PK. The distal tibial classic metaphyseal lesion: medial versus lateral cortical injury. *Pediatr Radiol* 2018;48:973-78.
37. Sieswerda-Hoogendoorn T, Boos S, Spivack B, Bilo RA, van Rijn RR. Abusive head trauma Part II: radiological aspects. *Eur J Pediatr* 2012;171:617-23.
38. Drubach LA, Johnston PR, Newton AW, Perez-Rossello JM, Grant FD, Kleinman PK. Skeletal trauma in child abuse: detection with 18F-NaF PET. *Radiology* 2010;255:173-81.
39. Bainbridge JK, Huey BM, Harrison SK. Should bone scintigraphy be used as a routine adjunct to skeletal survey in the imaging of non-accidental injury? A 10 year review of reports in a single centre. *Clin Radiol* 2015;70:e83-9.
40. Tsai A, McDonald AG, Rosenberg AE, Gupta R, Kleinman PK. High-resolution CT with histopathological correlates of the classic metaphyseal lesion of infant abuse. *Pediatr Radiol* 2014;44:124-40.
41. Cohen H. 3D Skull Reconstruction Aids for Diagnosing Child Abuse. *AAP Grand Rounds* 2013;30:43-43.
42. Orman G, Wagner MW, Seeburg D, et al. Pediatric skull fracture diagnosis: should 3D CT reconstructions be added as routine imaging? *J Neurosurg Pediatr* 2015;16:426-31.
43. Alanay Y, Lachman RS. A review of the principles of radiological assessment of skeletal dysplasias. *Journal of clinical research in pediatric endocrinology* 2011;3:163-78.
44. Veeramani AK, Higgins P, Butler S, et al. Diagnostic use of skeletal survey in suspected skeletal dysplasia. *Journal of clinical research in pediatric endocrinology* 2009;1:270-4.
45. Rimoin DL, Cohn D, Krakow D, Wilcox W, Lachman RS, Alanay Y. The skeletal dysplasias: clinical-molecular correlations. *Annals of the New York Academy of Sciences* 2007;1117:302-9.
46. Tuysuz B. A new concept of skeletal dysplasias. *The Turkish journal of pediatrics* 2004;46:197-203.
47. Chapman T, Sugar N, Done S, Marasigan J, Wambold N, Feldman K. Fractures in infants and toddlers with rickets. *Pediatr Radiol* 2010;40:1184-9.
48. Azouz EM, Saigal G, Rodriguez MM, Podda A. Langerhans' cell histiocytosis: pathology, imaging and treatment of skeletal involvement. *Pediatr Radiol* 2005;35:103-15.
49. McCarville MB. The child with bone pain: malignancies and mimickers. *Cancer imaging : the official publication of the International Cancer Imaging Society* 2009;9 Spec No A:S115-21.
50. Meyer JS, Nadel HR, Marina N, et al. Imaging guidelines for children with Ewing sarcoma and osteosarcoma: a report from the Children's Oncology Group Bone Tumor Committee. *Pediatric blood & cancer* 2008;51:163-70.
51. Ingram JD, Connell J, Hay TC, Strain JD, Mackenzie T. Oblique radiographs of the chest in nonaccidental trauma. *Emergency Radiology* 2000;7:42-46.
52. Jain N. The role of diagnostic imaging in the evaluation of child abuse. *British Columbia Medical Journal* 2015;57:336-40.
53. American College of Radiology. ACR practice parameter for communication of diagnostic imaging findings. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/CommunicationDiag.pdf>. Accessed January 14, 2020.
54. Barber I, Bixby SD, Morris NB, et al. An electronic tool for systematic reporting of fractures on skeletal surveys in suspected child abuse: prototype development and physician feedback. *Pediatr Radiol* 2014;44:1564-72.
55. Kleinman PL, Zurakowski D, Strauss KJ, et al. Detection of simulated inflicted metaphyseal fractures in a fetal pig model: image optimization and dose reduction with computed radiography. *Radiology* 2008;247:381-90.
56. Christian CW. The evaluation of suspected child physical abuse. *Pediatrics* 2015;135:e1337-54.
57. Jonsson A, Laurin S, Karner G, et al. Spatial resolution requirements in digital radiography of scaphoid fractures. An ROC analysis. *Acta Radiol* 1996;37:555-60.
58. Kottamasu SR, Kuhns LR, Stringer DA. Pediatric musculoskeletal computed radiography. *Pediatr Radiol* 1997;27:563-75.
59. Ludwig K, Schulke C, Diederich S, et al. Detection of subtle undisplaced rib fractures in a porcine model: radiation dose requirement--digital flat-panel versus screen-film and storage-phosphor systems. *Radiology* 2003;227:163-8.

60. Murphey MD, Bramble JM, Cook LT, Martin NL, Dwyer SJ, 3rd. Nondisplaced fractures: spatial resolution requirements for detection with digital skeletal imaging. *Radiology* 1990;174:865-70.
61. Offiah AC, Moon L, Hall CM, Todd-Pokropek A. Diagnostic accuracy of fracture detection in suspected non-accidental injury: the effect of edge enhancement and digital display on observer performance. *Clin Radiol* 2006;61:163-73.
62. Peer R, Lanser A, Giacomuzzi SM, et al. Storage phosphor radiography of wrist fractures: a subjective comparison of image quality at varying exposure levels. *Eur Radiol* 2002;12:1354-9.
63. Strotzer M, Gmeinwieser J, Spahn M, et al. Amorphous silicon, flat-panel, x-ray detector versus screen-film radiography: effect of dose reduction on the detectability of cortical bone defects and fractures. *Invest Radiol* 1998;33:33-8.
64. Wilson AJ, Mann FA, Murphy WA, Jr., Monsees BS, Linn MR. Photostimulable phosphor digital radiography of the extremities: diagnostic accuracy compared with conventional radiography. *AJR Am J Roentgenol* 1991;157:533-8.
65. Radiology ACo. ACR Endorses AAPM Position on Patient Gonadal and Fetal Shielding. Available at: <https://www.aapm.org/org/policies/details.asp?id=468&type=PP&current=true>. Accessed September 17, 2020.
66. American College of Radiology. ACR-AAPM technical standard for diagnostic medical physics performance monitoring of radiographic equipment. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/RadEquip.pdf>. Accessed January 14, 2020.

---

\*Practice parameters and technical standards are published annually with an effective date of October 1 in the year in which amended, revised or approved by the ACR Council. For practice parameters and technical standards published before 1999, the effective date was January 1 following the year in which the practice parameter or technical standard was amended, revised, or approved by the ACR Council.

#### Development Chronology for This Practice Parameter

- 1997 (Resolution 22)
- Revised 2001 (Resolution 31)
- Revised 2006 (Resolution 47, 17, 35)
- Amended 2009 (Resolution 11)
- Revised 2011 (Resolution 54)
- Amended 2014 (Resolution 39)
- Revised 2016 (Resolution 10)
- Revised 2021 (Resolution 37)