

American College of Radiology, with more than 30,000 members, is the principal organization of radiologists, radiation oncologists, and clinical medical physicists in the United States. The College is a nonprofit professional society whose primary purposes are to advance the science of radiology, improve radiologic services to the patient, study the socioeconomic aspects of the practice of radiology, and encourage continuing education for radiologists, radiation oncologists, medical physicists, and persons practicing in allied professional fields.

The American College of Radiology will periodically define new practice parameters and technical standards for radiologic practice to help advance the science of radiology and to improve the quality of service to patients throughout the United States. Existing practice parameters and technical standards will be reviewed for revision or renewal, as appropriate, on their fifth anniversary or sooner, if indicated.

Each practice parameter and technical standard, representing a policy statement by the College, has undergone a thorough consensus process in which it has been subjected to extensive review and approval. The practice parameters and technical standards recognize that the safe and effective use of diagnostic and therapeutic radiology requires specific training, skills, and techniques, as described in each document. Reproduction or modification of the published practice parameter and technical standard by those entities not providing these services is not authorized.

2022 (Resolution 29)\*

## **ACR–AIUM–SRU PRACTICE PARAMETER FOR THE PERFORMANCE OF PENILE ULTRASOUND**

---

### **PREAMBLE**

This document is an educational tool designed to assist practitioners in providing appropriate radiologic care for patients. Practice Parameters and Technical Standards are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care<sup>1</sup>. For these reasons and those set forth below, the American College of Radiology and our collaborating medical specialty societies caution against the use of these documents in litigation in which the clinical decisions of a practitioner are called into question.

The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the practitioner considering all the circumstances presented. Thus, an approach that differs from the guidance in this document, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in this document when, in the reasonable judgment of the practitioner, such course of action is indicated by variables such as the condition of the patient, limitations of available resources, or advances in knowledge or technology after publication of this document. However, a practitioner who employs an approach substantially different from the guidance in this document may consider documenting in the patient record information sufficient to explain the approach taken.

The practice of medicine involves the science, and the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment. Therefore, it should be recognized that adherence to the guidance in this document will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The purpose of this document is to assist practitioners in achieving this objective.

---

<sup>1</sup> *Iowa Medical Society and Iowa Society of Anesthesiologists v. Iowa Board of Nursing* 831 N.W.2d 826 (Iowa 2013) Iowa Supreme Court refuses to find that the *ACR Technical Standard for Management of the Use of Radiation in Fluoroscopic Procedures* (Revised 2008) sets a national standard for who may perform fluoroscopic procedures in light of the standard's stated purpose that ACR standards are educational tools and not intended to establish a legal standard of care. See also, *Stanley v. McCarver*, 63 P.3d 1076 (Ariz. App. 2003) where in a concurring opinion the Court stated that "published standards or guidelines of specialty medical organizations are useful in determining the duty owed or the standard of care applicable in a given situation" even though ACR standards themselves do not establish the standard of care.

## **I. INTRODUCTION**

The clinical aspects contained in specific sections of this practice parameter (Introduction, Indications, Specifications of the Examination, and Equipment Specifications) were developed collaboratively by the American College of Radiology (ACR), the American Institute of Ultrasound in Medicine (AIUM), and the Society of Radiologists in Ultrasound (SRU). Recommendations for physician requirements, written request for the examination, procedure documentation, and quality control vary between the 3 organizations and are addressed by each separately.

## **II. INDICATIONS**

Indications for penile ultrasound are based on the current literature recommendations and clinical practice standards.

1. Evaluation of signs and symptoms
  - a) Penile pain
  - b) Penile curvature
  - c) Erectile dysfunction
  - d) Penile mass
2. Suspected dorsal vein thrombosis
3. Evaluation of abnormal findings on physical examination of the phallus or urethra
4. Evaluation of penile tumors
5. Evaluation of a urethral stricture, diverticulum, or cyst
6. Evaluation of a calculus or foreign body of the phallus or urethra
7. Evaluation of penile trauma
8. Evaluation of priapism

## **III. QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL**

See the [ACR–SPR–SRU Practice Parameter for the Performance and Interpretation of Diagnostic Ultrasound Examinations](#) [1].

## **IV. SPECIFICATIONS OF THE EXAMINATION**

The written or electronic request for a penile ultrasound examination should provide sufficient information to demonstrate the medical necessity of the examination and allow for its proper performance and interpretation.

Documentation that satisfies medical necessity includes 1) signs and symptoms and/or 2) relevant history (including known diagnoses). Additional information regarding the specific reason for the examination or a provisional diagnosis would be helpful and may at times be needed to allow for the proper performance and interpretation of the examination.

The request for the examination must be originated by a physician or other appropriately licensed health care provider. The accompanying clinical information should be provided by a physician or other appropriately licensed health care provider familiar with the patient’s clinical problem or question and consistent with the state scope of practice requirements. (ACR Resolution 35 adopted in 2006 – revised in 2016, Resolution 12-b)

The phallus should be evaluated in at least 2 planes: longitudinal and transverse. Ventral and/or dorsal placement of the transducer should be used to obtain the best visualization of the corporal bodies and urethra [2,3]. Urethral examinations might require a sterile water-soluble intraurethral gel for detection of luminal and/or urethral wall pathology [4].

Transverse images should be obtained in the proximal, mid, and distal portions of the external portion of the phallus. Longitudinal views of the external portion of the phallus should be obtained of the right and left corpora cavernosa including the cavernosal artery. The non-external portions of the corpora cavernosa and urethra might be best visualized by perineal placement of the transducer.

The size and echogenicity of each corpus cavernosum should be compared to the contralateral side. Consider elastography for areas of unusual echogenicity [5]. If a palpable abnormality is the indication for the sonogram, this area should be directly imaged.

The evaluation of corporal vascular integrity requires the use of color and spectral Doppler imaging. Vascular assessment is performed by obtaining angle corrected velocities and spectral waveforms at discrete time intervals.

Evaluation for erectile dysfunction requires the use of color and spectral Doppler imaging before and after pharmacostimulation [3,6,7]. Use of pharmacostimulation should be done only after a discussion with the patient regarding the risks and benefits of the procedure and the potential for priapism. Patient is allowed to go home after 30 minutes of sonographic assessment when the diagnostic portion of the test is complete with clear instructions to return to the hospital/emergency room if erection persists for >4 hours and is painful.

For more specific protocol information for this procedure see reference [8].

## **V. DOCUMENTATION**

Reporting should be in accordance with the [ACR Practice Parameter for Communication of Diagnostic Imaging Findings](#) [9].

Adequate documentation is essential for high-quality patient care. There should be a permanent record of the ultrasound examination and its interpretation. Comparison with prior relevant imaging studies may prove helpful. Images of all appropriate areas, both normal and abnormal, should be recorded. Variations from normal size should generally be accompanied by measurements. Patient identification, facility identification, examination date, and image should be included in the DICOM header or displayed on the images. Images should indicate orientation. An official interpretation (final report) of the ultrasound examination should be included in the patient's medical record. Retention of the ultrasound examination images should be consistent both with clinical need and with relevant legal and local health care facility requirements.

## **VI. EQUIPMENT SPECIFICATIONS**

Equipment performance monitoring should be in accordance with the [ACR–AAPM Technical Standard for Diagnostic Medical Physics Performance Monitoring of Real Time Ultrasound Equipment](#) [10].

## **VII. QUALITY CONTROL AND IMPROVEMENT, SAFETY, INFECTION CONTROL, AND PATIENT EDUCATION**

Policies and procedures related to quality, patient education, infection control, and safety should be developed and implemented in accordance with the ACR Policy on Quality Control and Improvement, Safety, Infection Control, and Patient Education appearing under the heading *Position Statement on Quality Control & Improvement, Safety, Infection Control, and Patient Education* on the ACR website (<https://www.acr.org/Advocacy-and-Economics/ACR-Position-Statements/Quality-Control-and-Improvement>).

## **ACKNOWLEDGEMENTS**

This practice parameter was revised according to the process described under the heading *The Process for Developing ACR Practice Guidelines and Technical Standards* on the ACR website (<https://www.acr.org/Clinical-Resources/Practice-Parameters-and-Technical-Standards>) by the Committee on Practice Parameters – Ultrasound of the ACR Commission on Ultrasound and the Committee on Practice Parameters – Pediatric Radiology of the ACR Commissions on Pediatric Radiology in collaboration with the AIUM, and the SRU.

Writing Committee - members represent their societies in the initial and final revision of this practice parameter

ACR

Vikram S. Dogra, MD, Chair  
Christopher Fung, MD  
Akshya Gupta, MD  
Jamie Hui MD

AIUM

Bruce R. Gilbert, MD, PhD

SRU

Paul Hill, MD

Committee on Practice Parameters – Ultrasound

(ACR Committee responsible for sponsoring the draft through the process)

Sheila Sheth, MD, FACR, Chair  
Nirvikar Dahiya, MD, FAIUM, FSRU, Vice Chair  
Osama Ali, MD  
Marcela Böhm-Velez, MD, FACR  
Baljot S. Chahal, MD, MBA, BSc  
Christopher Fung, MD  
Helena Gabriel, MD  
Jamie Hui, MD

Stephen I. Johnson, MD  
Michelle L. Melany, MD, FACR  
Harriet J. Paltiel, MD  
Rupinder Penna, MD  
Kristin L. Rebik, DO  
Henrietta K. Rosenberg, MD, FACR  
Judy H. Squires, MD  
Joel P. Thompson, MD

Lauren P. Golding, MD, Chair, Commission on Ultrasound

David B. Larson, MD, MBA, Chair, Commission on Quality and Safety

Mary S. Newell, MD, FACR, Chair, Committee on Practice Parameters and Technical Standards

Comments Reconciliation Committee

Madelene Lewis, MD – CSC Chair  
Kurt Schoppe, MD– CSC Co-Chair  
Travis Glenn Browning, MD  
Timothy A. Crummy, MD, FACR  
Nirvikar Dahiya, MD, FAIUM, FSRU  
Vikram S. Dogra, MD  
Christopher Fung, MD  
Bruce R. Gilbert, MD, PhD  
Lauren P. Golding, MD

Akshya Gupta, MD  
Paul Hill, MD  
Jamie Hui MD  
Amy L. Kotsenas, MD, FACR  
David B. Larson, MD, MBA  
Mary S. Newell, MD, FACR  
Stacy D. O'Connor, MD  
William L. Simpson Jr, MD, FACR

**REFERENCES**

1. American College of Radiology. ACR–SPR–SRU Practice Parameter for the Performance and Interpretation of Diagnostic Ultrasound Examinations. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/US-Perf-Interpret.pdf>. Accessed December 4, 2020.
2. Herbener TE, Seftel AD, Nehra A, Goldstein I. Penile ultrasound. Seminars in urology 1994;12:320-32.
3. Shenoy-Bhangle A, Perez-Johnston R, Singh A. Penile imaging. Radiologic clinics of North America 2012;50:1167-81.
4. Shaida N, Berman LH. Ultrasound of the male anterior urethra. The British journal of radiology 2012;85 Spec No 1:S94-101.
5. Riversi V, Tallis V, Trovattelli S, et al. Realtime-elastosonography of the penis in patients with Peyronie's disease. Arch Ital Urol Androl 2012;84:174-7.

6. Halls J, Bydowell G, Patel U. Erectile dysfunction: the role of penile Doppler ultrasound in diagnosis. *Abdominal imaging* 2009;34:712-25.
7. Varela CG, Yeguas LAM, Rodríguez IC, Vila MDD. Penile Doppler Ultrasound for Erectile Dysfunction: Technique and Interpretation. *AJR. American journal of roentgenology* 2020;214:1112-21.
8. Ng A RG, Gilbert, BR. *Practical Urologic Ultrasound*. 3rd ed. New York, NY: Springer; 2021.
9. American College of Radiology. ACR Practice Parameter for Communication of Diagnostic Imaging Findings. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/CommunicationDiag.pdf>. Accessed December 4, 2020.
10. American College of Radiology. ACR–AAPM Technical Standard for Diagnostic Medical Physics Performance Monitoring of Real Time Ultrasound Equipment. Available at: <https://www.acr.org/-/media/ACR/Files/Practice-Parameters/US-Equip.pdf>. Accessed December 4, 2020.

---

\*Practice parameters and technical standards are published annually with an effective date of October 1 in the year in which amended, revised or approved by the ACR Council. For practice parameters and technical standards published before 1999, the effective date was January 1 following the year in which the practice parameter or technical standard was amended, revised, or approved by the ACR Council.

Development Chronology for this Practice Parameter

2022 (Resolution 29)

Amended 2023 (Resolution 2c)