

The American College of Radiology, with more than 30,000 members, is the principal organization of radiologists, radiation oncologists, and clinical medical physicists in the United States. The College is a nonprofit professional society whose primary purposes are to advance the science of radiology, improve radiologic services to the patient, study the socioeconomic aspects of the practice of radiology, and encourage continuing education for radiologists, radiation oncologists, medical physicists, and persons practicing in allied professional fields.

The American College of Radiology will periodically define new practice parameters and technical standards for radiologic practice to help advance the science of radiology and to improve the quality of service to patients throughout the United States. Existing practice parameters and technical standards will be reviewed for revision or renewal, as appropriate, on their fifth anniversary or sooner, if indicated.

Each practice parameter and technical standard, representing a policy statement by the College, has undergone a thorough consensus process in which it has been subjected to extensive review and approval. The practice parameters and technical standards recognize that the safe and effective use of diagnostic and therapeutic radiology requires specific training, skills, and techniques, as described in each document. Reproduction or modification of the published practice parameter and technical standard by those entities not providing these services is not authorized.

Revised 2014 (Resolution 2)\*

## **ACR–SAR–SCBT-MR PRACTICE PARAMETER FOR THE PERFORMANCE OF COMPUTED TOMOGRAPHY (CT) COLONOGRAPHY IN ADULTS**

---

### **PREAMBLE**

This document is an educational tool designed to assist practitioners in providing appropriate radiologic care for patients. Practice Parameters and Technical Standards are not inflexible rules or requirements of practice and are not intended, nor should they be used, to establish a legal standard of care<sup>1</sup>. For these reasons and those set forth below, the American College of Radiology and our collaborating medical specialty societies caution against the use of these documents in litigation in which the clinical decisions of a practitioner are called into question.

The ultimate judgment regarding the propriety of any specific procedure or course of action must be made by the practitioner in light of all the circumstances presented. Thus, an approach that differs from the guidance in this document, standing alone, does not necessarily imply that the approach was below the standard of care. To the contrary, a conscientious practitioner may responsibly adopt a course of action different from that set forth in this document when, in the reasonable judgment of the practitioner, such course of action is indicated by the condition of the patient, limitations of available resources, or advances in knowledge or technology subsequent to publication of this document. However, a practitioner who employs an approach substantially different from the guidance in this document is advised to document in the patient record information sufficient to explain the approach taken.

The practice of medicine involves not only the science, but also the art of dealing with the prevention, diagnosis, alleviation, and treatment of disease. The variety and complexity of human conditions make it impossible to always reach the most appropriate diagnosis or to predict with certainty a particular response to treatment. Therefore, it should be recognized that adherence to the guidance in this document will not assure an accurate diagnosis or a successful outcome. All that should be expected is that the practitioner will follow a reasonable course of action based on current knowledge, available resources, and the needs of the patient to deliver effective and safe medical care. The sole purpose of this document is to assist practitioners in achieving this objective.

---

<sup>1</sup> *Iowa Medical Society and Iowa Society of Anesthesiologists v. Iowa Board of Nursing*, \_\_ N.W.2d \_\_ (Iowa 2013) Iowa Supreme Court refuses to find that the *ACR Technical Standard for Management of the Use of Radiation in Fluoroscopic Procedures* (Revised 2008) sets a national standard for who may perform fluoroscopic procedures in light of the standard's stated purpose that ACR standards are educational tools and not intended to establish a legal standard of care. See also, *Stanley v. McCarver*, 63 P.3d 1076 (Ariz. App. 2003) where in a concurring opinion the Court stated that "published standards or guidelines of specialty medical organizations are useful in determining the duty owed or the standard of care applicable in a given situation" even though ACR standards themselves do not establish the standard of care.

## **I. INTRODUCTION**

Computed tomography colonography (CTC) is a minimally invasive structural examination of the colon and rectum to evaluate for colorectal polyps and neoplasms [1-13]. The goal of this examination is to establish the presence or absence of colorectal neoplasia by producing a diagnostic quality study at the lowest feasible radiation dose. This practice parameter outlines the performance of CTC in adult patients.

Individuals undergoing this examination may fall into one of several risk populations, and the examination may be designated as screening, surveillance, or diagnostic. There are several evidence-based guidelines that, with minor variations, categorize individuals into specific risk groups with correlated recommendations for management [14-17].

Screening identifies individuals who have colorectal cancer or adenomatous polyps without signs or symptoms of the disease. Individuals without other risk factors who are 50 years of age or older are at average risk. Individuals with a single first-degree relative (mother, father, sister, brother, or child) who have had colorectal neoplasia before age 60 or multiple first-degree relatives with colorectal neoplasia diagnosed at any age are defined as being at moderate risk. Average and moderate risk individuals are candidates for screening by CTC. Individuals with a longstanding history of inflammatory bowel disease or from families with defined genetic syndromes are at high risk and should not be considered for screening by CTC.

Surveillance involves the ongoing monitoring of people with previously diagnosed colorectal neoplasm identified as belonging to the high-risk category. The degree of risk may be related to the underlying or prior pathology. Surveillance CTC is also performed on individuals in whom colonic polyps have been previously identified but not resected, in order to assess stability of lesions that are considered low risk.

Diagnostic CTC examinations are performed on symptomatic individuals or as a follow-up to a prior but less definitive screening study. These individuals, by definition, are considered at greater risk of harboring colorectal neoplasia.

## **II. INDICATIONS AND CONTRAINDICATIONS**

### **A. Indications**

The indications for a CTC examination include, but are not limited to, the following:

1. Screening examination in individuals who are at average or moderate risk for developing colorectal carcinoma. Screening of individuals who are at moderate risk for colorectal cancer may be managed individually based on clinical context or local practice patterns.
2. Surveillance examination in patients with a history of previous colonic neoplasm [18], depending on the appropriate clinical context
3. Diagnostic examination in symptomatic patients, particularly in the setting of incomplete colonoscopy, including, but not limited to, those with the following:
  - a. Abdominal pain
  - b. Diarrhea
  - c. Constipation
  - d. Gastrointestinal bleeding
  - e. Anemia
  - f. Intestinal obstruction
  - g. Weight loss
4. Following incomplete screening, surveillance, or diagnostic colonoscopy and for characterization of colorectal lesions indeterminate on optical colonoscopy [19-23]
5. Patients who may be at increased risk for complications during optical colonoscopy (eg, advanced age, anticoagulant therapy, sedation risk, prior incomplete colonoscopy)

6. Follow-up of patients with a colonic stoma or after colectomy. Intubation of the stoma should be performed with caution to avoid colonic injury or perforation [24-26].
7. Prior to laparoscopic surgery for colorectal cancer in order to accurately localize the tumor or search for synchronous lesions

## B. Contraindications

1. The relative contraindications or conditions that require caution in performing a CTC examination include, but are not limited to, the following:
  - a. Symptomatic acute colitis
  - b. Acute diarrhea
  - c. Recent acute diverticulitis
  - d. Recent colorectal surgery
  - e. Symptomatic colon-containing abdominal wall hernia
  - f. Recent deep endoscopic biopsy or polypectomy/mucosectomy
  - g. Known or suspected colonic perforation
  - h. Symptomatic or high-grade small bowel obstruction
2. CTC is not indicated for the following:
  - a. Routine follow-up of inflammatory bowel disease
  - b. Hereditary polyposis or nonpolyposis cancer syndromes
  - c. Evaluation of anal canal disease
  - d. The pregnant or potentially pregnant patient (Refer to the [ACR–SPR Practice Parameter for Imaging Pregnant or Potentially Pregnant Adolescents and Women with Ionizing Radiation.](#))

## III. QUALIFICATIONS AND RESPONSIBILITIES OF PERSONNEL

See the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\).](#)

### A. Physician

The physician shall be responsible for all aspects of the study. The responsibilities include, but are not limited to, reviewing all indications for the examination; specifying and monitoring the appropriate patient preparation for colonic cleansing prior to the examination; specifying the appropriate imaging protocol, the methods of image reconstruction, and the use and dosage of contrast and pharmacologic agents; interpreting all resulting images and generating an official report; and assuring the quality of the images and the interpretation.

### Initial Training

1. For physicians with prior qualifications in general and/or abdominal-pelvic CT interpretation:

The radiologist or other physician who meets the qualifications of the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\)](#) will have substantial knowledge of radiation biology, the physics of CT scanning, the principles of CT image acquisition and postprocessing, including the use of diagnostic workstations, and the design of CT protocols, including rate and timing of contrast administration. The physician also will have substantial experience in CT interpretation, including CT of extracolonic structures that will be included on the CTC examination.

Supervising and interpreting physicians with prior qualifications in general and/or abdominal/pelvic CT interpretation shall also meet ONE of the following requirements:

- a. For physicians who receive their training in CTC in a training program approved by the Accreditation Council for Graduate Medical Education (ACGME), the Royal College of Physicians and Surgeons of Canada (RCPSC), the Collège des Médecins du Québec, or the American Osteopathic Association (AOA), such training shall include the following:
  - i. Education regarding patient preparation, bowel insufflation, and CT image acquisition.  
and
  - ii. Formal hands-on interactive training using dedicated CTC software, including the interpretation, reporting, and/or supervised review of at least 50 endoscopically confirmed CTC cases using primary 2-D and/or primary 3-D search with application of routine problem-solving techniques.

Ideally this collection of training cases will be chosen to demonstrate the gamut of appearances of colonic polyps and cancer and CTC interpretation pitfalls. Additionally, the cases should include examinations performed for a variety of indications (eg, screening, symptomatic, incomplete colonoscopy with subsequent validation) and acquisition techniques (eg, with and without fluid tagging and intravenous contrast).

or

- b. For physicians who receive their training in CTC after completing their residency or fellowship, such training shall include the following:
  - i. Education regarding patient preparation, bowel insufflation, and CT image acquisition.  
and
  - ii. Formal hands-on interactive training using dedicated CTC software, including the interpretation, reporting, and/or supervised review of at least 50 endoscopically confirmed CTC cases using primary 2D and/or primary 3D search employing commonly used problem-solving techniques.

Ideally this collection of training cases will be chosen to demonstrate the gamut of appearances of colonic polyps and cancer and CTC interpretation pitfalls. Additionally, the cases should include examinations performed for a variety of indications (eg, screening, symptomatic, incomplete colonoscopy with subsequent validation) and acquisition techniques (eg, with and without fluid tagging and intravenous contrast).

2. For physicians who do not have prior qualifications in general and/or abdominal-pelvic CT interpretation:

A radiologist or other physician who does not meet the qualifications of the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\)](#), or who meets these qualifications only for a specific anatomic area outside of the abdomen-pelvis, requires more extensive training and experience in CT scanning with an emphasis on the abdomen-pelvis and specific experience in CTC. In addition to specific training in imaging interpretation, this training must include knowledge of the principles of CT image acquisition and postprocessing, including the use of diagnostic workstations and the design of CT protocols, including rate and timing of contrast administration. The physician must also meet the same requirements, or document equivalent training, as those delineated in the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\)](#) with regard to knowledge of the physics of CT scanning and radiation biology. Some physicians will also require additional education in colon anatomy, physiology, and pathology.

Supervising and interpreting physicians without prior qualifications in general and/or abdominal-pelvic CT interpretation shall meet the following requirements:

- a. Completion of sufficient training and experience to meet the qualifications of the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\)](#). For a physician who assumes responsibilities for CT imaging exclusively in a specific anatomical area such as abdominal-pelvic CT and CTC, this includes the following:

- i. Completion of an ACGME-approved training program in their respective specialty in which they practice plus 200 hours of Category 1 CME in the performance and interpretation of abdominal-pelvic CT;  
and
  - ii. Supervision, interpretation, and reporting of 500 CT cases, at least 100 of which must be abdominal-pelvic CT cases during the past 36 months in a supervised situation;  
and
- b. Education regarding patient preparation, bowel insufflation, and CT image acquisition;  
and
  - c. Formal hands-on interactive training using dedicated CTC software, including the interpretation, reporting, and/or supervised review of at least 75 endoscopically confirmed CTC cases using primary 2-D and/or primary 3-D search with routine problem-solving techniques [27].

Ideally this collection of training cases will be chosen to demonstrate the gamut of appearances of colonic polyps and cancer and CTC interpretation pitfalls. Additionally, the cases should include examinations performed for a variety of indications (eg, screening, symptomatic, incomplete colonoscopy with subsequent validation) and acquisition techniques (eg, with and without fluid tagging and intravenous contrast).

#### Maintenance of Competence

When feasible, CTC training should be followed by a period of mentored supervision and double-reading by an experienced CTC-trained physician. A variety of other techniques may also be helpful for improving interpretive skills at CTC, including the following:

- Self-directed individual study of formal texts, atlases, review articles, and teaching files
- Testing with feedback
- Computer-aided detection algorithms, which can be used as a second reader

A total of 50 cases every 2 years should be reviewed to maintain skills in CTC. This can be accomplished in several ways, such as:

- Performance of CTC in local practice, with follow-up of positive findings with endoscopy or surgery
- CME-sponsored reviews online, DVDs, or at review courses where case interpretation precedes disclosure of the correct answers

#### B. Radiologic Technologist

Qualifications of the radiologic technologist should include familiarity with the technical requirements of performing CTC, including selection of scanning parameters, rectal tip insertion, proper patient positioning, colonic insufflation of room air and carbon dioxide with manual and automated techniques, tube removal, and quality assurance of the exam prior to discharge of the patient.

### IV. SPECIFICATIONS OF THE EXAMINATION

The written or electronic request for CT colonography should provide sufficient information to demonstrate the medical necessity of the examination and allow for its proper performance and interpretation.

Documentation that satisfies medical necessity includes 1) signs and symptoms and/or 2) relevant history (including known diagnoses). Additional information regarding the specific reason for the examination or a provisional diagnosis would be helpful and may at times be needed to allow for the proper performance and interpretation of the examination.

The request for the examination must be originated by a physician or other appropriately licensed health care provider. The accompanying clinical information should be provided by a physician or other appropriately licensed health care provider familiar with the patient's clinical problem or question and consistent with the state scope of practice requirements. (ACR Resolution 35, adopted in 2006)

#### A. Colon Preparation

Preparation of the colon for CTC should consist of a combination of dietary restriction, hydration, saline cathartics, and/or low-volume polyethylene glycol agents and contact laxatives. The intent is to achieve a colon that is free of fecal material and excess fluid or as close to this ideal as possible [28-31] .

When feasible, the use of tagging is recommended unless the patient refuses or cannot comply. The use of water-soluble contrast alone or in combination with low-volume barium is the preferred tagging method. The goal of tagging is to passively incorporate contrast into any residual fluid and stool in order to raise their inherent CT densities, which helps to discriminate these residua from the soft-tissue density of polyps or advanced cancers. Noncathartic or reduced-cathartic approaches to CTC bowel preparation (also known as “prepress” or “minimal prep” CTC) aim to reduce patient discomfort associated with pre-examination bowel purgation. Although data supporting the success of this approach continue to emerge, a fully cleansed colon is generally recommended for patients who can comply [32-35]. For patients who cannot comply with a standard preparation or who are too fragile to undergo a standard preparation, sufficient data exists to justify limited cathartic CTC when combined with a tagging agent [13,36-38].

#### B. Examination Technique

1. The medical history, including patient compliance with the colon preparation, should be reviewed.
2. The patient should evacuate prior to insertion of the rectal tube. A soft (nonrigid) tip tube is recommended.
3. The rectal tube tip should be inserted by a physician or a trained assistant (radiologic technologist, nurse, or physician assistant). If a rectal retention balloon is used, inflation should be discontinued if the patient complains of persistent severe pain. This may indicate an increased risk of perforation.
4. If a rectal retention balloon is used, it should be deflated or advanced on one series to facilitate detection of low lying rectal lesions.
5. The use of antispasmodics is not considered necessary for routine examination, and the evidence for improved distention or patient comfort remains inconclusive. No benefit is seen with glucagon [39,40]. There may be some benefit with hyoscine N-butylbromide [41,42]. However, this agent is currently unavailable in the United States.
6. The preferred method of colonic insufflation is by means of mechanical insufflation using carbon dioxide [43]; however, manual insufflation with room air is acceptable. A sufficient volume of carbon dioxide or room air should be administered either with an automatic insufflator or manually to provide full colon distention [44].
7. The adequacy of colon distention should be checked with a localizer to ensure a complete and full column of gas throughout the colon before each CT series acquisition.
8. Complete anatomic imaging of the colon and rectum should be obtained in at least 2 patient positions (usually supine and prone) [30,45]. If the patient is unable to tolerate prone positioning, a lateral decubitus can be substituted. Each series should be obtained in end expiration to minimize pressure effects of inflated lungs on the transverse colon.
9. Screening studies should be performed using a low-dose, nonenhanced CT technique on a multidetector CT scanner [3,46-49]. Colonography studies should be performed such that there is appropriate adaptation of  $CTDI_{vol}$  to patient size, using either technique charts or automatic exposure control. The recommended

radiation output for routine screening CTC, quantified using  $CTDI_{vol}$ , should be less than or equal to one-half the diagnostic reference level for routine abdominal pelvic CT (2008 ACR CT Accreditation Program) or one-quarter of this value per position (ie,  $CTDI_{vol}$  of 6.25 mGy per position or a total of 12.5 mGy for dual-position CTC). Generally, for scans performed at a tube potential of 120 kVp, this requires an effective mAs value between 50 and 80 (where effective mAs is equal to the tube current-time product (mAs) divided by the spiral pitch value). Because these factors may not be appropriate for every CT scanner model, the scan protocol parameters should be adjusted as necessary to obtain the required image quality at or below the suggested  $CTDI_{vol}$  values (6.25 mGy per scan position or 12.5 mGy total for the supine and prone position scans).

10. The use of dose reduction techniques is encouraged. These include reductions in tube current (mA), exposure time(s), tube current-time product (mAs) or tube potential (kV). Automatic exposure control systems, image-based noise reduction algorithms, and iterative reconstruction techniques can also be used to effectively reduce dose [50-52]. Using these strategies, much lower radiation doses for screening CTC can be achieved, similar to the average annual background level of radiation in the range of 3 mSv or less [53].
11. Additional imaging after repositioning and reinsufflation may be needed to adequately distend a colonic segment. Additional imaging (eg, in right or left decubitus position) is appropriate when imaging in 2 positions fails to adequately display the colonic lumen and acquisition of additional data is likely to result in a diagnostic study [54]. Any additional imaging should be limited to the segment of interest in order to minimize additional radiation dose.
12. For morbidly obese patients, radiation dose should be appropriately increased to maintain diagnostic image quality [55].
13. Diagnostic CTC examinations should use the same CT parameters as screening CTC examinations. Diagnostic CTC may occasionally require intravenous (IV) contrast to characterize intracolonic or extracolonic structures or to address a second medical indication. When IV contrast is used, the dose on the contrast-enhanced series should be similar to a standard abdominal pelvic CT; the supine series is typically used for this. Thus for diagnostic contrast-enhanced CT colonography, a typical order of sequences should include an initial low dose noncontrast prone series followed by a supine series with intravenous contrast and normal dose.
14. CTC is optimally performed on a multidetector ( $\geq 16$  slice) CT scanner. A section thickness of 1–1.25 mm with a reconstruction interval of  $\leq 1$  mm is optimal. The breathhold should not exceed 25 seconds.
15. Networking capability should be available to transfer the image data to a workstation with specialized software for CTC interpretation.

### C. Quality Control

The following quality controls should be applied to all CTC examinations:

1. Colon cleansing and distention should be adequate for detecting polyps 1 cm or larger, at a minimum.
2. There should be complete anatomic coverage of the colon and rectum.
3. The luminal surface of each segment of the colon should be visualized in at least 1 position. Suboptimally visualized colonic segments should be reimaged. The use of lateral decubitus views or reinsufflation may be helpful in cases of suboptimal distention or excessive fluid [54].
4. Efforts should be made to ensure a diagnostic quality examination before the patient leaves the facility.
5. The following is suggested for a quality control program:
  - a. Radiologic, endoscopic, and pathologic findings should be correlated whenever available.
  - b. Detection rates for colorectal cancer and polyps of 1 cm or greater should be determined and periodically monitored. There should be an assessment of false-positive rates for all reported polyps in patients who undergo subsequent colonoscopy.
  - c. Participation in the ACR NRDR<sup>®</sup> CTC Registry (<https://nrdr.acr.org/Portal/CTC/Main/page.aspx>) is recommended, with regular comparison of facility data to national data to determine how local detection and complication rates compare with national rates, and whether performance is adequate or if further internal review is indicated. Please note, the American Board of Radiology has deemed that the CTC

Registry meets the criteria for practice quality improvement toward the purpose of fulfilling requirements in the ABR Maintenance of Certification Program.

#### D. Data Interpretation

The purpose of CTC is to accurately evaluate the colon for the presence or absence of clinically significant neoplastic lesions. Abnormalities may range from discrete mucosal elevations (which may be malignant or at risk to become malignant) to infiltrating tumors. Lesion morphology (sessile, pedunculated, flat, mass) and segmental location should be reported.

1. Detection and characterization of colorectal findings

Workstations utilized for CTC interpretation should be able to display 2-D and 3-D data as well as prone and supine data side-by-side for interactive interrogation. The software should also allow the interpreting physician to change the window width and level settings interactively and in real time.

If an abnormality is suspected during either primary 2-D or 3-D search:

- a. The abnormality should be interrogated with multiplanar reconstruction (MPR) and multiple endoluminal views to evaluate the morphology of the suspected lesion.
  - b. Supine and prone data should be evaluated to determine if the lesion is mobile. Causes of mobility include residual fecal material, pedunculated polyp, or a rotating colon segment. Most true polyps can be identified in both the supine and prone views; potential lesions seen on only one view have a much lower predictive value.
  - c. The window setting should be adjusted between colon and soft-tissue settings to determine if the lesion shows homogeneous soft-tissue attenuation or is heterogeneous.
2. Measurement of colorectal findings  
Polyps should be measured using optimized multiplanar reconstruction (ie, axial, sagittal, or coronal view, which best elongates lesion) and/or 3-D images. Measurement of the size of the lesion should be based on the largest diameter of the polyp head (excluding stalk if present) or at the base of a sessile lesion [57,58].
  3. Extracolonic findings

Extracolonic structures should be evaluated at the time of the review of the colon. Significant or potentially significant abnormalities should be included in the report. A study optimized for evaluating colon abnormalities may not be optimal for detecting and characterizing extracolonic abnormalities. Specifically, detecting incidental findings with low subject contrast may be limited with aggressive dose reduction on unenhanced images. This limitation is reduced somewhat by increasing the section thickness for the extracolonic reconstruction (eg, 5-mm thick sections at 3-mm intervals), which reduces noise and decreases the number of images that need to be reviewed for incidental lesion detection. Abnormalities or questionable abnormalities in structures unrelated to the colon may be identified during the process of reviewing the 2-D axial images of the colon. A balanced approach for recommending further workup of extracolonic findings should weigh the likelihood of a clinically important finding against the increased cost, patient anxiety, and potential complications related to additional evaluation [59].

#### V. DOCUMENTATION AND COMMUNICATION OF RESULTS

Reporting should be in accordance with the [ACR Practice Parameter for Communication of Diagnostic Imaging Findings](#).

Any colonic segment that cannot be adequately evaluated for technical reasons should be documented as such. Polyps  $\geq 6$  mm should be identified and reported. Consistent with the 2008 American Cancer Society recommendations [60], these patients should be offered polypectomy at colonoscopy, understanding that clinical management may vary depending on the patient's age, risk to undergo colonoscopy, other significant comorbidities, or the preference of the patient or the referring physician. Recommendations for clinical management options may be incorporated into the report.

In patients with only diminutive polyps  $\leq 5$  mm, the risk of high-grade dysplasia or cancer is extremely low [6,10,61]. Although there continues to be debate about patients with only diminutive polyps, the clinical risk of these diminutive polyps is extremely small [48,62,63]. The benefits of polypectomy versus 5-year surveillance need to be balanced with the broader risks, including the costs and complications of polypectomy. Namely, given the low risk of advanced neoplasia along with the low specificity of diminutive lesions at CTC, a large number of patients could be referred to endoscopy inappropriately [64]. Furthermore at colonoscopy, concern over decreased productivity for false-positive CTC examinations has been raised [65], in addition to the low rate of detecting small lesions at colonoscopy [66,67]. Current CTC acquisition techniques targeted at the index lesion size of  $\geq 6$  mm with low-dose techniques do not always optimize detection of diminutive lesions. Given these considerations, the ACR currently does not believe that reporting of these diminutive lesions is necessary [59].

Extracolonic abnormalities of potential medical significance should also be reported. Good patient care mandates that CTC interpretation include full evaluation of the numerous extracolonic structures and that findings of potential clinical significance be reported and communicated in a clear and timely fashion. However, most extracolonic findings are not clinically significant in screening/asymptomatic cohorts. In screening cohorts, the prevalence of clinically significant extracolonic findings is low [62,68-72]. Caution should be used in the interpretation and reporting of findings likely to be of low clinical significance in order to avoid unnecessary subsequent/serial diagnostic examinations and associated patient anxiety [59].

Clarity and consistency of reporting the colonic and extracolonic findings are critical for effective implementation. There is increasing use of the CT Colonography Reporting and Data System (C-RADS), which is a consensus statement of a standardized reporting structure for CTC findings published in 2005, modeled after the Breast Imaging Reporting and Data System<sup>®</sup> (BI-RADS<sup>®</sup>) reporting of mammography [59]. The reporting structure of C-RADS describes how to report lesion size, morphology, and location with a summary category score per patient.

## **VI. EQUIPMENT SPECIFICATIONS**

Examinations should be performed with MDCT ( $\geq 16$  slice) equipment meeting all applicable federal and state radiation standards. The CT scanner should have the capability of providing section thicknesses of 1–1.25 mm and reconstruction intervals  $\leq 1$  mm at breathholds of less than 25 seconds. Equipment should provide diagnostic image quality and networking capability.

## **VII. RADIATION SAFETY IN IMAGING**

Radiologists, medical physicists, registered radiologist assistants, radiologic technologists, and all supervising physicians have a responsibility for safety in the workplace by keeping radiation exposure to staff, and to society as a whole, “as low as reasonably achievable” (ALARA) and to assure that radiation doses to individual patients are appropriate, taking into account the possible risk from radiation exposure and the diagnostic image quality necessary to achieve the clinical objective. All personnel that work with ionizing radiation must understand the key principles of occupational and public radiation protection (justification, optimization of protection and application of dose limits) and the principles of proper management of radiation dose to patients (justification, optimization and the use of dose reference levels) [http://www-pub.iaea.org/MTCD/Publications/PDF/Pub1578\\_web-57265295.pdf](http://www-pub.iaea.org/MTCD/Publications/PDF/Pub1578_web-57265295.pdf)

Nationally developed guidelines, such as the ACR’s [Appropriateness Criteria<sup>®</sup>](#), should be used to help choose the most appropriate imaging procedures to prevent unwarranted radiation exposure.

Facilities should have and adhere to policies and procedures that require varying ionizing radiation examination protocols (plain radiography, fluoroscopy, interventional radiology, CT) to take into account patient body habitus (such as patient dimensions, weight, or body mass index) to optimize the relationship between minimal radiation dose and adequate image quality. Automated dose reduction technologies available on imaging equipment should be used whenever appropriate. If such technology is not available, appropriate manual techniques should be used.

Additional information regarding patient radiation safety in imaging is available at the Image Gently® for children ([www.imagegently.org](http://www.imagegently.org)) and Image Wisely® for adults ([www.imagewisely.org](http://www.imagewisely.org)) websites. These advocacy and awareness campaigns provide free educational materials for all stakeholders involved in imaging (patients, technologists, referring providers, medical physicists, and radiologists).

Radiation exposures or other dose indices should be measured and patient radiation dose estimated for representative examinations and types of patients by a Qualified Medical Physicist in accordance with the applicable ACR Technical Standards. Regular auditing of patient dose indices should be performed by comparing the facility's dose information with national benchmarks, such as the ACR Dose Index Registry, the NCRP Report No. 172, Reference Levels and Achievable Doses in Medical and Dental Imaging: Recommendations for the United States or the Conference of Radiation Control Program Director's National Evaluation of X-ray Trends. (ACR Resolution 17 adopted in 2006 – revised in 2009, 2013, Resolution 52).

## **VIII. QUALITY CONTROL AND IMPROVEMENT, SAFETY, INFECTION CONTROL, AND PATIENT EDUCATION**

Policies and procedures related to quality, patient education, infection control, and safety should be developed and implemented in accordance with the ACR Policy on Quality Control and Improvement, Safety, Infection Control, and Patient Education appearing under the heading *Position Statement on QC & Improvement, Safety, Infection Control, and Patient Education* on the ACR website (<http://www.acr.org/guidelines>).

For specific issues regarding CT quality control, see the [ACR Practice Parameter for Performing and Interpreting Diagnostic Computed Tomography \(CT\)](#).

Equipment performance monitoring should be in accordance with the [ACR–AAPM Technical Standard for Medical Physics Performance Monitoring of Computed Tomography \(CT\) Equipment](#).

## **ACKNOWLEDGEMENTS**

This practice parameter was revised according to the process described under the heading *The Process for Developing ACR Practice Parameters and Technical Standards* on the ACR website (<http://www.acr.org/guidelines>) by the ACR Colon Cancer Committee of the Commission on General, Small, and Rural Practice.

### Collaborative Committee

Members represent their societies in the initial and final revision of this practice parameter.

#### ACR

Judy Yee, MD, FACR, Chair  
Abraham H. Dachman, MD, FACR  
David H. Kim, MD, FACR  
Perry J. Pickhardt, MD  
Jonathan D. Stephenson, MD  
Michael Zalis, MD

#### SAR

Kevin J. Chang, MD  
W. Ross Stevens, MD, FACR

## SCBT-MR

Elizabeth G. McFarland, MD, FACR

## Colon Cancer Committee

Judy Yee, MD, FACR, Chair

Matthew A. Barish, MD, FACR

Abraham H. Dachman, MD, FACR

J. G. Fletcher, MD

C. Daniel Johnson, MD, FACR

David H. Kim, MD, FACR

Mark E. Klein, MD

Paul M. Knechtges, MD

Elizabeth G. McFarland, MD, FACR

Perry J. Pickhardt, MD

Michael Zalis, MD

Cynthia H. McCollough, PhD, FACR, FAAPM, Consultant

James A. Brink, MD, FACR, Chair, Commission on Body Imaging

Debra L. Monticciolo, MD, FACR, Chair, Commission on Quality and Safety

Julie K. Timins, MD, FACR, Chair, Committee on Practice Parameters and Technical Standards

## Comments Reconciliation Committee

Lawrence R. Muroff, MD, FACR, Chair

Joseph G. Cernigliaro, MD, FACR, Co-Chair

Kimberly E. Applegate, MD, MS, FACR

James A. Brink, MD, FACR

Kevin J. Chang, MD

Dianna D. Cody, PhD

Abraham H. Dachman, MD, FACR

William T. Herrington, MD, FACR

David H. Kim, MD, FACR

Elizabeth G. McFarland, MD, FACR

Debra L. Monticciolo, MD, FACR

Perry J. Pickhardt, MD

Jonathan D. Stephenson, MD

W. Ross Stevens, MD, FACR

Julie K. Timins, MD, FACR

Judy Yee, MD, FACR

Michael Zalis, MD

## **REFERENCES**

1. Fenlon HM, Nunes DP, Schroy PC, 3rd, Barish MA, Clarke PD, Ferrucci JT. A comparison of virtual and conventional colonoscopy for the detection of colorectal polyps. *N Engl J Med*. 1999;341(20):1496-1503.
2. Graser A, Stieber P, Nagel D, et al. Comparison of CT colonography, colonoscopy, sigmoidoscopy and faecal occult blood tests for the detection of advanced adenoma in an average risk population. *Gut*. 2009;58(2):241-248.
3. Halligan S, Altman DG, Taylor SA, et al. CT colonography in the detection of colorectal polyps and cancer: systematic review, meta-analysis, and proposed minimum data set for study level reporting. *Radiology*. 2005;237(3):893-904.
4. Johnson CD, Chen MH, Toledano AY, et al. Accuracy of CT colonography for detection of large adenomas and cancers. *N Engl J Med*. 2008;359(12):1207-1217.

5. Johnson CD, Harmsen WS, Wilson LA, et al. Prospective blinded evaluation of computed tomographic colonography for screen detection of colorectal polyps. *Gastroenterology*. 2003;125(2):311-319.
6. Kim DH, Pickhardt PJ, Taylor AJ, et al. CT colonography versus colonoscopy for the detection of advanced neoplasia. *N Engl J Med*. 2007;357(14):1403-1412.
7. Laghi A, Iannaccone R, Carbone I, et al. Detection of colorectal lesions with virtual computed tomographic colonography. *Am J Surg*. 2002;183(2):124-131.
8. Macari M, Bini EJ, Xue X, et al. Colorectal neoplasms: prospective comparison of thin-section low-dose multi-detector row CT colonography and conventional colonoscopy for detection. *Radiology*. 2002;224(2):383-392.
9. Mulhall BP, Veerappan GR, Jackson JL. Meta-analysis: computed tomographic colonography. *Ann Intern Med*. 2005;142(8):635-650.
10. Pickhardt PJ, Choi JR, Hwang I, et al. Computed tomographic virtual colonoscopy to screen for colorectal neoplasia in asymptomatic adults. *N Engl J Med*. 2003;349(23):2191-2200.
11. Pickhardt PJ, Nugent PA, Mysliwiec PA, Choi JR, Schindler WR. Location of adenomas missed by optical colonoscopy. *Ann Intern Med*. 2004;141(5):352-359.
12. Yee J, Akerkar GA, Hung RK, Steinauer-Gebauer AM, Wall SD, McQuaid KR. Colorectal neoplasia: performance characteristics of CT colonography for detection in 300 patients. *Radiology*. 2001;219(3):685-692.
13. Zalis ME, Blake MA, Cai W, et al. Diagnostic accuracy of laxative-free computed tomographic colonography for detection of adenomatous polyps in asymptomatic adults: a prospective evaluation. *Ann Intern Med*. 2012;156(10):692-702.
14. Winawer S, Fletcher R, Rex D, et al. Colorectal cancer screening and surveillance: clinical guidelines and rationale-Update based on new evidence. *Gastroenterology*. 2003;124(2):544-560.
15. Winawer SJ, Fletcher RH, Miller L, et al. Colorectal cancer screening: clinical guidelines and rationale. *Gastroenterology*. 1997;112(2):594-642.
16. Lieberman D. Colorectal cancer screening: practice guidelines. *Dig Dis*. 2012;30 Suppl 2:34-38.
17. Rex DK, Johnson DA, Anderson JC, Schoenfeld PS, Burke CA, Inadomi JM. American College of Gastroenterology guidelines for colorectal cancer screening 2009 [corrected]. *Am J Gastroenterol*. 2009;104(3):739-750.
18. Kim HJ, Park SH, Pickhardt PJ, et al. CT colonography for combined colonic and extracolonic surveillance after curative resection of colorectal cancer. *Radiology*. 2010;257(3):697-704.
19. Copel L, Sosna J, Kruskal JB, Raptopoulos V, Farrell RJ, Morrin MM. CT colonography in 546 patients with incomplete colonoscopy. *Radiology*. 2007;244(2):471-478.
20. Fenlon HM, McAneny DB, Nunes DP, Clarke PD, Ferrucci JT. Occlusive colon carcinoma: virtual colonoscopy in the preoperative evaluation of the proximal colon. *Radiology*. 1999;210(2):423-428.
21. Macari M, Berman P, Dicker M, Milano A, Megibow AJ. Usefulness of CT colonography in patients with incomplete colonoscopy. *AJR Am J Roentgenol*. 1999;173(3):561-564.
22. Morrin MM, Farrell RJ, Raptopoulos V, McGee JB, Bleday R, Kruskal JB. Role of virtual computed tomographic colonography in patients with colorectal cancers and obstructing colorectal lesions. *Dis Colon Rectum*. 2000;43(3):303-311.
23. Neri E, Giusti P, Battolla L, et al. Colorectal cancer: role of CT colonography in preoperative evaluation after incomplete colonoscopy. *Radiology*. 2002;223(3):615-619.
24. Laghi A, Rengo M, Graser A, Iafrate F. Current status on performance of CT colonography and clinical indications. *European journal of radiology*. 2012.
25. Lee JH, Park SH, Lee SS, et al. CT colonography in patients who have undergone sigmoid colostomy: a feasibility study. *AJR Am J Roentgenol*. 2011;197(4):W653-657.
26. Leonardou P, Striggaris K, Pappas P, et al. Screening of patients after colectomy: virtual colonography. *Abdominal imaging*. 2006;31(5):521-528.
27. Rockey DC, Barish M, Brill JV, et al. Standards for gastroenterologists for performing and interpreting diagnostic computed tomographic colonography. *Gastroenterology*. 2007;133(3):1005-1024.
28. Kim DH, Pickhardt PJ, Hinshaw JL, Taylor AJ, Mukherjee R, Pfau PR. Prospective blinded trial comparing 45-mL and 90-mL doses of oral sodium phosphate for bowel preparation before computed tomographic colonography. *J Comput Assist Tomogr*. 2007;31(1):53-58.
29. Macari M, Lavelle M, Pedrosa I, et al. Effect of different bowel preparations on residual fluid at CT colonography. *Radiology*. 2001;218(1):274-277.

30. Yee J, Kumar NN, Hung RK, Akerkar GA, Kumar PR, Wall SD. Comparison of supine and prone scanning separately and in combination at CT colonography. *Radiology*. 2003;226(3):653-661.
31. Borden ZS, Pickhardt PJ, Kim DH, Lubner MG, Agriantonis DJ, Hinshaw JL. Bowel preparation for CT colonography: blinded comparison of magnesium citrate and sodium phosphate for catharsis. *Radiology*. 2010;254(1):138-144.
32. Iannaccone R, Laghi A, Catalano C, et al. Computed tomographic colonography without cathartic preparation for the detection of colorectal polyps. *Gastroenterology*. 2004;127(5):1300-1311.
33. Lefere P, Gryspeerdt S, Marrannes J, Baekelandt M, Van Holsbeeck B. CT colonography after fecal tagging with a reduced cathartic cleansing and a reduced volume of barium. *AJR Am J Roentgenol*. 2005;184(6):1836-1842.
34. Zalis ME, Perumpillichira JJ, Magee C, Kohlberg G, Hahn PF. Tagging-based, electronically cleansed CT colonography: evaluation of patient comfort and image readability. *Radiology*. 2006;239(1):149-159.
35. Fletcher JG, Johnson C.D, Welch T.J., et al. Optimization of CT colonography technique: prospective trial in 180 patients. *Radiology*. 2000;216:704-711.
36. Florie J, van Gelder RE, Schutter MP, et al. Feasibility study of computed tomography colonography using limited bowel preparation at normal and low-dose levels study. *Eur Radiol*. 2007;17(12):3112-3122.
37. Jensch S, Bipat S, Peringa J, et al. CT colonography with limited bowel preparation: prospective assessment of patient experience and preference in comparison to optical colonoscopy with cathartic bowel preparation. *Eur Radiol*. 2010;20(1):146-156.
38. Keedy AW, Yee J, Aslam R, et al. Reduced cathartic bowel preparation for CT colonography: prospective comparison of 2-L polyethylene glycol and magnesium citrate. *Radiology*. 2011;261(1):156-164.
39. Morrin MM, Farrell RJ, Keogan MT, Kruskal JB, Yam CS, Raptopoulos V. CT colonography: colonic distention improved by dual positioning but not intravenous glucagon. *Eur Radiol*. 2002;12(3):525-530.
40. Yee J, Hung RK, Akerkar GA, Wall SD. The usefulness of glucagon hydrochloride for colonic distention in CT colonography. *AJR Am J Roentgenol*. 1999;173(1):169-172.
41. de Haan MC, Boellaard TN, Bossuyt PM, Stoker J. Colon distension, perceived burden and side-effects of CT-colonography for screening using hyoscine butylbromide or glucagon hydrochloride as bowel relaxant. *European journal of radiology*. 2012;81(8):e910-916.
42. Rogalla P, Lembcke A, Ruckert JC, et al. Spasmolysis at CT colonography: butyl scopolamine versus glucagon. *Radiology*. 2005;236(1):184-188.
43. Shinnars TJ, Pickhardt PJ, Taylor AJ, Jones DA, Olsen CH. Patient-controlled room air insufflation versus automated carbon dioxide delivery for CT colonography. *AJR Am J Roentgenol*. 2006;186(6):1491-1496.
44. Burling D, Taylor SA, Halligan S, et al. Automated insufflation of carbon dioxide for MDCT colonography: distension and patient experience compared with manual insufflation. *AJR Am J Roentgenol*. 2006;186(1):96-103.
45. Chen SC, Lu DS, Hecht JR, Kadell BM. CT colonography: value of scanning in both the supine and prone positions. *AJR Am J Roentgenol*. 1999;172(3):595-599.
46. Iannaccone R, Laghi A, Catalano C, et al. Detection of colorectal lesions: lower-dose multi-detector row helical CT colonography compared with conventional colonoscopy. *Radiology*. 2003;229(3):775-781.
47. van Gelder RE, Venema HW, Florie J, et al. CT colonography: feasibility of substantial dose reduction--comparison of medium to very low doses in identical patients. *Radiology*. 2004;232(2):611-620.
48. van Gelder RE, Venema HW, Serlie IW, et al. CT colonography at different radiation dose levels: feasibility of dose reduction. *Radiology*. 2002;224(1):25-33.
49. Ginsburg M, Obara P, Wise L, Wroblewski K, Vannier MW, Dachman AH. BMI-based radiation dose reduction in CT colonography. *Academic radiology*. 2013;20(4):486-492.
50. Chang KJ, Yee J. Dose reduction methods for CT colonography. *Abdominal imaging*. 2013;38(2):224-232.
51. Flicek KT, Hara AK, Silva AC, Wu Q, Peter MB, Johnson CD. Reducing the radiation dose for CT colonography using adaptive statistical iterative reconstruction: A pilot study. *AJR Am J Roentgenol*. 2010;195(1):126-131.
52. Yoon MA, Kim SH, Lee JM, et al. Adaptive statistical iterative reconstruction and Veo: assessment of image quality and diagnostic performance in CT colonography at various radiation doses. *J Comput Assist Tomogr*. 2012;36(5):596-601.
53. Boellaard TN, Venema HW, Streekstra GJ, Stoker J. Effective radiation dose in CT colonography: is there a downward trend? *Academic radiology*. 2012;19(9):1127-1133.

54. Buchach CM, Kim DH, Pickhardt PJ. Performing an additional decubitus series at CT colonography. *Abdominal imaging*. 2011;36(5):538-544.
55. McCollough CH, Bruesewitz MR, Kofler JM, Jr. CT dose reduction and dose management tools: overview of available options. *Radiographics*. 2006;26(2):503-512.
56. Schmidt B, Kalender WA. A fast voxel-based Monte Carlo method for scanner and patient specific dose calculations in computed tomography. *Physica Medica*. 2002;18(3):43-53.
57. McFarland EG, Brink JA, Pilgram TK, et al. Spiral CT colonography: reader agreement and diagnostic performance with two- and three-dimensional image-display techniques. *Radiology*. 2001;218(2):375-383.
58. Pickhardt PJ, Lee AD, McFarland EG, Taylor AJ. Linear polyp measurement at CT colonography: in vitro and in vivo comparison of two-dimensional and three-dimensional displays. *Radiology*. 2005;236(3):872-878.
59. Zalis ME, Barish MA, Choi JR, et al. CT colonography reporting and data system: a consensus proposal. *Radiology*. 2005;236(1):3-9.
60. Levin B, Lieberman DA, McFarland B, et al. Screening and surveillance for the early detection of colorectal cancer and adenomatous polyps, 2008: a joint guideline from the American Cancer Society, the US Multi-Society Task Force on Colorectal Cancer, and the American College of Radiology. *CA Cancer J Clin*. 2008;58(3):130-160.
61. Lieberman D, Moravec M, Holub J, Michaels L, Eisen G. Polyp Size and Advanced Histology in Patients Undergoing Colonoscopy Screening: Implications for CT Colonography. *Gastroenterology*. 2008.
62. Flicker MS, Tsoukas AT, Hazra A, Dachman AH. Economic impact of extracolonic findings at computed tomographic colonography. *J Comput Assist Tomogr*. 2008;32(4):497-503.
63. Lenhart DK, Zalis ME. Debate: diminutive polyps noted at CT colonography need not be reported. *Gastrointestinal endoscopy clinics of North America*. 2010;20(2):227-237.
64. Kim DH, Pooler BD, Weiss JM, Pickhardt PJ. Five year colorectal cancer outcomes in a large negative CT colonography screening cohort. *Eur Radiol*. 2012;22(7):1488-1494.
65. Hur C, Gazelle GS, Hsu EH, Halpern EF, Podolsky DK. The effect of prior colonic imaging on endoscopic productivity: potential impact of computed tomographic colonography. *Clin Gastroenterol Hepatol*. 2005;3(11):1124-1127.
66. Hixson LJ, Fennerty MB, Sampliner RE, McGee D, Garewal H. Prospective study of the frequency and size distribution of polyps missed by colonoscopy. *J Natl Cancer Inst*. 1990;82(22):1769-1772.
67. Rex DK, Cutler CS, Lemmel GT, et al. Colonoscopic miss rates of adenomas determined by back-to-back colonoscopies. *Gastroenterology*. 1997;112(1):24-28.
68. Chin M, Mendelson R, Edwards J, Foster N, Forbes G. Computed tomographic colonography: prevalence, nature, and clinical significance of extracolonic findings in a community screening program. *Am J Gastroenterol*. 2005;100(12):2771-2776.
69. Gluecker TM, Johnson CD, Wilson LA, et al. Extracolonic findings at CT colonography: evaluation of prevalence and cost in a screening population. *Gastroenterology*. 2003;124(4):911-916.
70. Pickhardt PJ, Kim DH, Taylor AJ, Burnside ES. CT colonography reporting and data system (C-RADS): prospective categorization for screening in 2,501 patients. Paper presented at: 2006 RSNA Scientific Assembly 2006.
71. Pickhardt PJ, Hanson ME, Vanness DJ, et al. Unsuspected extracolonic findings at screening CT colonography: clinical and economic impact. *Radiology*. 2008;249(1):151-159.
72. Yee J, Kumar NN, Godara S, et al. Extracolonic abnormalities discovered incidentally at CT colonography in a male population. *Radiology*. 2005;236(2):519-526.

---

\*Practice parameters and technical standards are published annually with an effective date of October 1 in the year in which amended, revised or approved by the ACR Council. For practice parameters and technical standards published before 1999, the effective date was January 1 following the year in which the practice parameter or technical standard was amended, revised, or approved by the ACR Council.

#### Development Chronology for this Practice Parameter

2005 (Resolution 29)

Amended 2006 (Resolution 17, 35)

Revised 2009 (Resolution 36)

Revised 2014 (Resolution 2)